



Schedule of Scope to Certificate of Conformity Approved Process

IECQ Certificate No.: IECQ-L ULTW 09.0029

CB Certificate No.: T1142

Schedule Number: IECQ-L ULTW 09.0029-S Rev No.: 1 Revision Date: 2018/10/15 Page 1 of 1

Appendix-1 (T1142) Schedule of Scope to Certificate of Approval

Calibration	Range	Standard
Spectral irradiance	200nm to 2500nm	CIE 63:1984 The spectro-radiometric measurement of light sources
Averaged LED intensity (ILED A and ILED B)	360nm to 1100nm	CIE 127:2007 Measurement of LEDs
Luminous flux of LEDs	360nm to 1100nm	CIE 127:2007 Measurement of LEDs
Luminous flux and spectral radiant flux	360nm to 1100nm	CIE 63:1984 The spectro-radiometric measurement of light sources

Technical Reviewer of DQS: *Michael Chou* Date: 10/25/2018

This schedule is only valid in conjunction with the referenced Certificate of Approval
This approval and any schedule(s) may only be reproduced in full.
This approval is not transferable and remains the property of the issuing body.
The Status and authenticity of this approval and any schedule(s) may be verified by visiting the
Official IECQ Website. www.iecq.org

DQS-Group - DQS Taiwan Inc., Feng Yuan Dist., Taichung City, Taiwan

