



IEC QUALITY ASSESSMENT SYSTEM (IECQ)

For rules and details of the IECQ visit www.iecq.org

IECQ Verification Statement Carbon Footprint of Product

IECQ Certificate No.: IECQ-V LCIE 25.0003

CB Reference No.: 21934008

Annex Number: IECQ-V LCIE 25.0003-A

Rev No.: 1

Revision Date: 2025/10/30

Page 1 of 1

Product information:

The product system includes:

- | | | |
|---|---------------------------|-------------------|
| - | 2 x Surgical lights | (2xARD570002921) |
| - | 2 x Spring arms | (2xARD568921931) |
| - | 1 x Suspension | (1xARD569720930) |
| - | 1 x Anchoring tube | (1xARD569063999) |
| - | 1 x Anchoring flange | (1xARD568604902) |
| - | 1 x Power supply | (1xARD567355951) |
| - | 18 x Sterilizable handles | (18xARD567917900) |

Functional Unit Details:

Function provided: Lighting the operating field with two cupolas

Extent of the function: 2x 1 meter distance and light spot diameter higher than 15cm

Expected level of quality: 2x 100 000 lux at 1 meter distance

Duration/lifetime of the product: 10 years, 10 hours a day, 300 days a year

This annex is only valid in conjunction with the referenced Verification Statement
This statement and any schedule(s) may only be reproduced in full.
This statement is not transferable and remains the property of the issuing IECQ CB.
The Status and authenticity of this statement may be verified by visiting www.iecq.org

LCIE - 33 avenue du Général Leclerc 92260 FONTENAY-AUX-ROSES - France



LCIE