



**Schedule of Scope to Certificate of Approval**  
**Independent Testing Laboratory**  
 IECQ Certificate No.: IECQ-L ULTW 09.0016  
 CB Certificate No.: T1093

Schedule Number: IECQ-L ULTW 09.0016-S      Rev No.: 4      Revision Date: 2016/05/24      Page 1 of 1

Calibration Items	Calibration Method	Ranges	Best measurement Capability
DC volt meter	MTI-T0079-14	100 mV	0.11%
		1 V	0.006%
		10 V	0.005%
		100 V	0.007%
		1000 V	0.007%
DC amper meter	MTI-T0079-13	1 mA	0.29%
		10 mA	0.08%
		100 mA	0.06%
		1 A	0.13%
AC volt meter	MTI-T0079-12	1 V	0.11%
		10 V	0.11%
		100 V	0.11%
		700 V	0.12%
		(1 KHz)	
AC amper meter	MTI-T0079-11	100 mA	0.58%
		1 A	0.21%
		(1 kHz)	
Ohm meter (Null below)	MTI-T0079-10	1 Ω	0.48%
		10 Ω	0.06%
		100 Ω	0.02%
		10 KΩ	0.013%
		10 MΩ	0.05%

Remarks : Best measurement capability represents an expanded uncertainty using a coverage factor approximately 95% level of confidence.

This schedule is only valid in conjunction with the referenced Certificate of Approval  
 This approval and any schedule(s) may only be reproduced in full.  
 This approval is not transferable and remains the property of the issuing body.  
 The Status and authenticity of this approval and any schedule(s) may be verified by visiting the  
 Official IECQ Website. [www.iecq.org](http://www.iecq.org)



**DQS Group-DQS Taiwan Inc., Feng Yuan Dist., Taichung City, Taiwan**