



## Schedule of Scope to Certificate of Approval Independent Testing Laboratory

IECQ Certificate No.: IECQ-L ULTW 09.0013

CB Certificate No.: 20004501 ITL

Schedule Number: IECQ-L ULTW 09.0013-S Rev No.: 14 Revision Date: 2025/10/06 Page 1 of 1

### Appendix-1 (20004501 ITL) Schedule of Scope to Certificate of Approval

Calibration Items	Standard/Method
RTDs	CL3002 / RTD -80 ~ 400°C
Thermocouples	CL3003 , CL3004 , CL3016 / Sensor -80 ~1554°C
	CL3014 / Instrument -200 ~ 1750°C
	CL3015 / Furnace -150 ~ 1750°C
Infrared thermometers	CL3020 / Sensor 35 ~ 500°C
Thermal hygrometers	CL3022 / Sensor 10 ~ 30°C & 30 ~ 80%
	CL3023 / Instrument 10 ~ 30°C & 30 ~ 80%
	CL3024 / Furnace 10 ~ 30°C & 30 ~ 80%
Pressure transducers.	CL3026 / 0 ~ 250 PSI
Field Test	AMS2750 /CL3015&CL3106 / Furnace Instrument – 150 ~ 1750°C
Timer	AMS2750 /CL3015 / Instrument 1 ~ 28800 seconds

Technical Reviewer of DQS: Paul Yang Date: 10/06/2025

This schedule is only valid in conjunction with the referenced Certificate of Approval  
This approval and any schedule(s) may only be reproduced in full.  
This approval is not transferable and remains the property of the issuing body.  
The Status and authenticity of this approval and any schedule(s) may be verified by visiting the  
Official IECQ Website. [www.iecq.org](http://www.iecq.org)

DQS-Group - DQS Taiwan Inc., Feng Yuan Dist., Taichung City, Taiwan

